

Predetermined Change Control Plans for Medical Devices

Draft Guidance for Industry and Food and Drug Administration Staff

DRAFT GUIDANCE

This draft guidance document is being distributed for comment purposes only.

Document issued on August 22, 2024.

You should submit comments and suggestions regarding this draft document within 90 days of publication in the *Federal Register* of the notice announcing the availability of the draft guidance. Submit electronic comments to <https://www.regulations.gov>. Submit written comments to the Dockets Management Staff, Food and Drug Administration, 5630 Fishers Lane, Room 1061, (HFA-305), Rockville, MD 20852-1740. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*.

For questions about this document regarding CDRH-regulated devices, contact the Office of Product Evaluation and Quality (OPEQ), Regulation, Policy, and Guidance staff at RPG@fda.hhs.gov. For questions about this document regarding CBER-regulated devices, contact the Office of Communication, Outreach, and Development (OCOD) at 1-800-835-4709 or 240-402-8010, or by email at ocod@fda.hhs.gov.



U.S. Department of Health and Human Services
Food and Drug Administration
Center for Devices and Radiological Health
Center for Biologics Evaluation and Research

Preface

40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57

Additional Copies

CDRH

Additional copies are available from the Internet. You may also send an email request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please include the document number GUI00007026 and complete title of the guidance in the request.

CBER

Additional copies are available from the Center for Biologics Evaluation and Research (CBER), Office of Communication, Outreach, and Development (OCOD), 10903 New Hampshire Ave., Bldg. 71, Room 3128, Silver Spring, MD 20993-0002, or by calling 1-800-835-4709 or 240-402-8010, by email, ocod@fda.hhs.gov, or from the Internet at <https://www.fda.gov/vaccines-blood-biologics/guidance-compliance-regulatory-information-biologics/biologics-guidances>.

Table of Contents

58		
59		
60		
61	I.	Introduction..... 1
62	II.	Background..... 2
63	III.	Scope..... 4
64	IV.	Guiding Principles for PCCPs 6
65	V.	Policy for PCCPs 8
66	A.	Components of a PCCP 9
67	B.	Establishing a PCCP..... 9
68	C.	Identifying a PCCP in a Marketing Submission 11
69	D.	Utilizing an Authorized PCCP to Implement Device Modifications..... 13
70	E.	Modifying a Previously Authorized PCCP 16
71	F.	Version Control and Maintenance of a PCCP for a Device..... 17
72	VI.	Types of Modifications 19
73	A.	Determining Whether a Modification may be Appropriate for Inclusion in a PCCP in a 74 510(k) or De Novo Submission 20
75	B.	Determining Whether a Modification may be Appropriate for Inclusion in a PCCP in a 76 PMA Application or Supplement..... 22
77	VII.	Recommended Content for a PCCP..... 25
78	A.	Description of Modifications 25
79	(1)	Goals of the Description of Modifications..... 26
80	(2)	Content of the Description of Modifications 26
81	B.	Modification Protocol 26
82	(1)	Goals of the Modification Protocol..... 27
83	(2)	Content of the Modification Protocol Section 28
84	C.	Traceability Between the Description of Modifications Section and the Modification 85 Protocol Section 30
86	D.	Impact Assessment..... 30
87	VIII.	Examples of Modifications for PCCPs 31
88	IX.	Sample of 510(k) Summary Information Regarding the PCCP..... 37
89		

Predetermined Change Control Plans for Medical Devices

Draft Guidance for Industry and Food and Drug Administration Staff

This draft guidance, when finalized, will represent the current thinking of the Food and Drug Administration (FDA or Agency) on this topic. It does not establish any rights for any person and is not binding on FDA or the public. You can use an alternative approach if it satisfies the requirements of the applicable statutes and regulations. To discuss an alternative approach, contact the FDA staff or Office responsible for this guidance as listed on the title page.

I. Introduction

On December 29, 2022, section 3308 of the Food and Drug Omnibus Reform Act of 2022, Title III of Division FF of the Consolidated Appropriations Act, 2023, Pub. L. No. 117-328 (“FDORA”) added section 515C “Predetermined Change Control Plans for Devices” to the Federal Food, Drug, and Cosmetic (FD&C) Act. Section 515C of the FD&C Act (21 U.S.C. 360e-4) has provisions regarding predetermined change control plans (PCCPs) for devices requiring premarket approval (PMA) or premarket notification (510(k)). A PCCP is the documentation describing what modifications will be made to a device and how the modifications will be assessed. This draft guidance provides FDA’s current thinking on a policy for PCCPs and recommendations on the information to include in a PCCP in a marketing submission for a device. This draft guidance recommends that a PCCP describe the planned device modifications, the associated methodology to develop, validate, and implement those modifications, and an assessment of the impact of those modifications. FDA reviews the PCCP as part of a marketing submission for a device to ensure the continued safety and effectiveness of the device without necessitating additional marketing submissions for implementing each modification described in the PCCP.

Upon finalization of this guidance, we will make conforming Level 2 updates to the following guidance documents to clarify that a PMA supplement or 510(k) is not required for a change to a device consistent with an approved or cleared PCCP¹:

¹ Section 515C(a)(1) and 515C(b)(1) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- [“Modifications to Devices Subject to Premarket Approval \(PMA\) - The PMA Supplement Decision-Making Process”](#)
- [“Deciding When to Submit a 510\(k\) for a Change to an Existing Device”](#)
- [“Deciding When to Submit a 510\(k\) for a Software Change to an Existing Device”](#)

Upon finalization of this guidance, we may also make Level 2 updates to the [“Marketing Submission Recommendations for a Predetermined Change Control Plan for Artificial Intelligence/Machine Learning \(AI/ML\)-Enabled Device Software Functions”](#) guidance document and other device-specific guidance documents containing information on PCCPs for consistency with this guidance.

This guidance has been prepared by the Center for Devices and Radiological Health (CDRH) and the Center for Biologics Evaluation and Research (CBER) in consultation with the Center for Drug Evaluation and Research (CDER) and the Office of Combination Products (OCP) in the Office of the Commissioner.

In general, FDA’s guidance documents do not establish legally enforceable responsibilities. Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidances means that something is suggested or recommended, but not required.

II. Background

The concept of a PCCP is not entirely new to FDA. For example, in 2017, FDA described in the [“Deciding When to Submit a 510\(k\) for a Change to an Existing Device”](#) guidance that changes in the expiration date for use of a device generally do not require submission of a new 510(k) when the same methods or protocols that are described in the previously cleared 510(k) are used to support the change. In 2022, FDA described in the [“Replacement Reagent and Instrument Family Policy for In Vitro Diagnostic Devices”](#) guidance how manufacturers may add certain additional instruments for use with an in vitro diagnostic (IVD) assay that was previously cleared for use with a specific instrument without submission of a new 510(k), in part, by conducting a risk-based assessment and design verification and/or validation activities to assess the use of the IVD assay with the new instrument(s).

In 2019, FDA introduced the term and description of a PCCP in the [“Proposed Regulatory Framework for Modifications to Artificial Intelligence/Machine Learning \(AI/ML\)-Based Software as a Medical Device \(SaMD\) - Discussion Paper and Request for Feedback.”](#)² This discussion paper described a potential approach to premarket review of PCCPs for AI/ML-based software modifications. On December 29, 2022, FDORA was enacted. Section 3308 of this legislation, titled “Predetermined Change Control Plans for Devices,” amended the FD&C Act to add section 515C, which has provisions regarding PCCPs for devices that would otherwise

² Also available at FDA’s website on [“Artificial Intelligence and Machine Learning in Software as a Medical Device.”](#)

Contains Nonbinding Recommendations

Draft – Not for Implementation

165 require a PMA supplement or a new 510(k). In 2023, we issued a draft guidance titled
166 [“Marketing Submission Recommendations for a Predetermined Change Control Plan for](#)
167 [Artificial Intelligence/Machine Learning \(AI/ML\)-Enabled Device Software Functions,”](#) which
168 incorporated stakeholder feedback on the discussion paper and reflected our initial thinking on
169 the statutory change and the types of information we recommend be included in a PCCP in a
170 marketing submission. In that draft guidance, we proposed a policy to support iterative
171 improvement through prospective FDA review and authorization of modifications to an AI/ML-
172 enabled device software function while continuing to provide a reasonable assurance of device
173 safety and effectiveness. That draft guidance provided recommendations on the information to
174 include in a PCCP in a marketing submission for an AI/ML-enabled device software function.
175 We also proposed that a PCCP should include information on the following topics that could be
176 divided into three categories: Description of Modifications (a detailed description of the specific,
177 planned device modifications), Modification Protocol (the associated methodology to develop,
178 validate, and implement those modifications), and Impact Assessment (the assessment of the
179 benefits and risks of implementing a PCCP and the plan for risk mitigation).

180
181 In other device-specific guidance, FDA has also recommended using PCCPs to implement
182 specific modifications for certain device types. For example, later in 2023, we issued a final
183 guidance titled [“Antimicrobial Susceptibility Test \(AST\) System Devices – Updating](#)
184 [Breakpoints in Device Labeling”](#) describing how manufacturers may use PCCPs to update
185 susceptibility test interpretative criteria and associated performance in device labeling in
186 response to changes posted on the [FDA-Recognized Antimicrobial Susceptibility Test](#)
187 [Interpretive Criteria](#) website.

188
189 This broader draft guidance focuses on section 515C(a)-(c) of the FD&C Act, reproduced below,
190 and provides FDA’s policy and recommendations for marketing submissions for PCCPs for all
191 device types.

SECTION 515C. [21 U.S.C. 360e-4] PREDETERMINED CHANGE CONTROL PLANS FOR DEVICES.

(a) APPROVED DEVICES.—

195 (1) IN GENERAL.—Notwithstanding section 515(d)(5)(A), a supplemental application shall not
196 be required for a change to a device approved under section 515, if such change is consistent with
197 a predetermined change control plan that is approved pursuant to paragraph (2).

198 (2) PREDETERMINED CHANGE CONTROL PLAN.—The Secretary may approve a
199 predetermined change control plan submitted in an application, including a supplemental
200 application, under section 515 that describes planned changes that may be made to the device (and
201 that would otherwise require a supplemental application under section 515), if the device remains
202 safe and effective without any change.

203 (3) SCOPE.—The Secretary may require that a change control plan include labeling required for
204 safe and effective use of the device as such device changes pursuant to such plan, notification
205 requirements if the device does not function as intended pursuant to such plan, and performance
206 requirements for changes made under the plan.

(b) CLEARED DEVICES.—

208 (1) IN GENERAL.—Notwithstanding section 510(k), a premarket notification shall not be
209 required for a change to a device cleared under section 510(k), if such change is consistent with an
210 established predetermined change control plan granted pursuant to paragraph (2).

211 (2) PREDETERMINED CHANGE CONTROL PLAN.—The Secretary may clear a
212 predetermined change control plan submitted in a notification submitted under section 510(k) that

Contains Nonbinding Recommendations

Draft – Not for Implementation

213 describes planned changes that may be made to the device (and that would otherwise require a
214 new notification), if—
215 (A) the device remains safe and effective without any such change; and
216 (B) the device would remain substantially equivalent to the predicate.
217 (3) SCOPE.—The Secretary may require that a change control plan include labeling required for
218 safe and effective use of the device as such device changes pursuant to such plan, notification
219 requirements if the device does not function as intended pursuant to such plan, and performance
220 requirements for changes made under the plan.
221 (c) PREDICATE DEVICES.—In making a determination of substantial equivalence pursuant to section
222 513(i), the Secretary shall not compare a device to changed versions of a device implemented in accordance
223 with an established predetermined change control plan as a predicate device. Only the version of the device
224 cleared or approved, prior to changes made under the predetermined change control plan, may be used by a
225 sponsor as a predicate device.
226

227 **III. Scope**

228 This draft guidance proposes recommendations on the types of modifications that, at this time,
229 FDA believes generally may be appropriate for inclusion in a PCCP, and the information that
230 should be included in a PCCP in a marketing submission³ for any device type.
231

232 For purposes of this guidance, a PCCP includes those device modifications that generally would
233 otherwise require a new marketing submission.^{4,5} These modifications include those that could
234 significantly affect,⁶ or that otherwise affect,⁷ the safety or effectiveness of the device,⁸ unless
235 those modifications are covered by an authorized PCCP. By including a PCCP in a marketing
236 submission for a device, manufacturers can prospectively specify and seek premarket
237 authorization⁹ for intended modifications to a device without needing to submit additional
238 marketing submissions or obtain further FDA authorization before implementing such
239 modifications – provided the changes are implemented consistent with the PCCP that has been

³ For purposes of this guidance, the term “marketing submission” includes a PMA application, 510(k) submission, or De Novo Classification request.

⁴ For purposes of this guidance, unless otherwise stated, references to “device modifications” or “modifications” are those that generally would otherwise require a new marketing submission pursuant to 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

⁵ For more information on whether a modification would require a new marketing submission, see the FDA guidances “[Modifications to Devices Subject to Premarket Approval \(PMA\) – The PMA Supplement Decision-Making Process](#),” “[Deciding When to Submit a 510\(k\) for a Change to an Existing Device](#),” or “[Deciding When to Submit a 510\(k\) for a Software Change to an Existing Device](#),” referred to as the “Device Modifications guidances” hereafter.

⁶ 21 CFR 807.81(a)(3).

⁷ 21 CFR 814.39(a).

⁸ In accordance with 21 CFR 807.81(a)(3), a 510(k) is required for significant changes or modifications to a device and include 1) those that “could significantly affect the safety or effectiveness of the device, e.g., a significant change or modification in design, material, chemical composition, energy source, or manufacturing process” or include 2) “a major change or modification in the intended use of the device.” In accordance with 21 CFR 814.39(a), a PMA supplement is required for “change[s] affecting the safety or effectiveness of the device” unless an exception applies (see 21 CFR 814.39). For simplicity, in this guidance, the phrase “significant changes” or “significant modifications” refers to 21 CFR 807.81(a)(3). However, for devices subject to PMA requirements, the broader requirement pursuant to 21 CFR 814.39(a) of a “change affecting the safety or effectiveness” applies.

⁹ For purposes of this guidance, the term “authorization” includes approval of a PMA application, clearance of a 510(k) submission, or grant of a De Novo Classification request.

Contains Nonbinding Recommendations

Draft – Not for Implementation

240 reviewed and established through a device marketing authorization (referred to hereafter as the
241 “authorized PCCP”).¹⁰ In other words, obtaining FDA authorization of a PCCP as part of a
242 marketing submission for a device allows a manufacturer to modify its device over time in
243 accordance with the PCCP instead of obtaining separate FDA authorization for each significant
244 change prior to each implementation.¹¹

245
246 Because a PCCP is appropriate for device modifications that generally would otherwise require a
247 new marketing submission,¹² this guidance does not address a plan that only contains a proposal
248 for modifications that would not require a new marketing submission. For such modifications,
249 the Quality System regulation (QSR) (21 CFR Part 820)¹³ requires manufacturers of finished
250 medical devices to, among other things, document the change in the device master record.¹⁴ For
251 devices subject to PMA requirements, such modifications need to be reported to FDA in post-
252 approval periodic reports required as a condition to approval of the device.¹⁵

253
254 Premarket authorization for a device with a PCCP may be established through the PMA pathway
255 (see section 515C(a) of the FD&C Act), the 510(k) pathway (see section 515C(b) of the FD&C
256 Act), or the De Novo pathway (see section 513(f)(2) of the FD&C Act).¹⁶ For devices subject to
257 510(k) requirements, in making a determination of substantial equivalence where the predicate
258 device was authorized with a PCCP, the subject device must be compared to the version of the
259 predicate device cleared or approved prior to changes made under the PCCP.¹⁷

260

¹⁰ For purposes of this guidance, the term “authorized PCCP” refers to a PCCP that has been reviewed and established through a device marketing authorization.

¹¹ Sections 515C(a)(1) and 515C(b)(1) of the FD&C Act, 21 CFR 807.81(a)(3), and 21 CFR 814.39(a).

¹² 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

¹³ On February 2, 2024, FDA issued a final rule amending the device QSR, 21 CFR Part 820, to align more closely with international consensus standards for devices ([89 FR 7496](#)). This final rule will take effect on February 2, 2026. Once in effect, this rule will withdraw the majority of the current requirements in Part 820 and instead incorporate by reference the 2016 edition of the International Organization for Standardization (ISO) 13485, Medical devices – Quality management systems – Requirements for regulatory purposes, in Part 820. As stated in the final rule, the requirements in ISO 13485 are, when taken in totality, substantially similar to the requirements of the current Part 820, providing a similar level of assurance in a firm’s quality management system and ability to consistently manufacture devices that are safe and effective and otherwise in compliance with the FD&C Act. When the final rule takes effect, FDA will also update the references to provisions in 21 CFR Part 820 in this guidance to be consistent with that rule.

¹⁴ 21 CFR 820.181.

¹⁵ See 21 CFR 814.39(b) and 21 CFR 814.82(a)(7) and FDA’s guidance “[Annual Reports for Approved Premarket Approval Applications \(PMA\)](#).”

¹⁶ The De Novo classification process allows FDA to classify a device into class I or II when general controls or general controls and special controls provide a reasonable assurance of safety and effectiveness, but for which there is no legally marketed predicate. The De Novo pathway, therefore, allows FDA to develop special controls that provide a reasonable assurance of the safety and effectiveness of the subject device. At this time, FDA expects that if it authorizes a device with a PCCP via the De Novo pathway, the Agency would develop appropriate special controls, which may include specific requirements for a PCCP.

¹⁷ See section 515C(c) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

261 Generally, the recommendations in this guidance apply to the device constituent part¹⁸ of device-
262 led¹⁹ combination products.²⁰ The recommendations in this guidance do not apply to the drug or
263 biologic constituent part of device-led combination products. For device-led combination
264 products with a PCCP, the FDA review division will consult CBER or CDER, as appropriate.
265

266 FDA highly encourages early engagement regarding a proposed PCCP with the FDA review
267 division; in particular, early engagement could be especially helpful for certain devices,
268 including combination products and high-risk, life-sustaining, life-supporting, or implantable
269 devices. FDA encourages manufacturers to leverage the Q-Submission Program²¹ for obtaining
270 FDA feedback on a proposed PCCP for a device prior to submitting a marketing submission. The
271 FDA review division with purview over the device will provide feedback on a proposed PCCP,
272 including whether the scope of the modifications is appropriate for inclusion in a PCCP and,
273 based on the statutory and regulatory requirements applicable to that device, what evidence and
274 information would ensure that the device under that PCCP remains safe and effective under
275 section 515C of the FD&C Act.
276

277 This draft guidance is intended to provide recommendations on the information to include in a
278 PCCP in a marketing submission for a device. This draft guidance is not intended to provide a
279 complete description of what may be necessary to include in a marketing submission for a
280 device.²² The proposed recommendations in this draft guidance do not change applicable
281 statutory and regulatory standards, such as device clearance or approval standards, nor the
282 applicable requirements, including marketing submission content requirements and the
283 requirements for valid scientific evidence.²³ FDA recommends that you refer to other guidances,
284 as applicable to a specific device, for recommendations on aspects of the submission beyond the
285 PCCP.
286

287 This draft guidance is intended to provide recommendations on the types of modifications that, at
288 this time, FDA believes generally may be appropriate for inclusion in a PCCP for a device. This
289 draft guidance is not intended to delineate a comprehensive list of modifications FDA would
290 consider appropriate for inclusion in a PCCP for a device. However, the draft guidance proposes
291 types of modifications that generally may be appropriate (see Section VI. of this guidance) and
292 provides illustrative examples (see Section VIII. of this guidance).
293

294 **IV. Guiding Principles for PCCPs**

295 In developing this guidance for manufacturers and FDA staff on PCCPs for all device types,
296 several guiding principles were followed. Some derive from existing policies and others are key
297 to understanding the policy proposed in this draft guidance. Anyone using this draft guidance

¹⁸ See 21 CFR 4.2.

¹⁹ See 21 CFR 3.4 for information on lead center assignment.

²⁰ See 21 CFR 3.2(e).

²¹ See FDA's guidance "[Requests for Feedback and Meetings for Medical Device Submissions: The Q-Submission Program](#)," hereafter referred to as the "Q-Submission Program."

²² See, e.g., 21 CFR 807.87, 21 CFR 860.220, or 21 CFR 814.20.

²³ See, e.g., 21 CFR 807.87, 21 CFR 860.220, or 21 CFR 814.20, and 21 CFR 860.7(c).

Contains Nonbinding Recommendations

Draft – Not for Implementation

298 should bear in mind the following guiding principles:
299

- 300
- 301 • **Reasonable assurance of safety and effectiveness and substantial equivalence of**
302 **devices with PCCPs** – A PCCP is part of the device marketing authorization. As such,
303 for the PCCP to be authorized with the device, the totality of the information included in
304 a PCCP should enable FDA to assess the reasonable assurance of safety and effectiveness
305 or substantial equivalence of the device.²⁴

 - 306 • **PCCPs may be a least burdensome option to support device modifications** –
307 Manufacturers may wish to use PCCPs as a way to implement modifications to their
308 devices without needing to submit a new marketing submission for each modification
309 while continuing to provide a reasonable assurance of device safety and effectiveness.
310 When used appropriately, PCCPs are expected to be least burdensome²⁵ for
311 manufacturers and FDA. However, PCCPs are optional. FDA will review the device and
312 PCCP and determine the acceptability of a proposed PCCP in accordance with applicable
313 device approval or clearance standards under the FD&C Act and its implementing
314 regulations.²⁶ FDA’s review of the device and PCCP will follow a risk-based approach
315 with consideration of the device’s intended use and technological characteristics, as well
316 as the regulatory history of the specific device, device type, and manufacturer, and use
317 FDA’s benefit-risk framework.²⁷

 - 318
 - 319 • **PCCPs are part of a device’s marketing authorization** – A PCCP is part of the
320 device’s marketing authorization, and as such, the manufacturer is required to implement
321 modifications consistent with their authorized PCCP, when the manufacturer chooses to
322 implement those modifications and use the PCCP to do so.²⁸ Premarket authorization of a
323 PCCP is based on the details of the specific PCCP developed by the manufacturer for that
324 specific device. When a manufacturer includes a PCCP in a marketing submission for a
325 device, the FDA review division with purview over the device will determine the
326 acceptability of a proposed PCCP. This will include whether the scope of the
327 modifications is appropriate for inclusion in a PCCP and, based on the statutory and
328 regulatory requirements applicable to that device, what information and evidence would
329 ensure that the device under that PCCP remains safe and effective under section 515C of
330 the FD&C Act. When the PCCP is authorized, the PCCP is a part of the marketing
331 authorization of a device and will be included in the device’s letter of authorization.
332

 - 333 • **PCCPs are specific** – A PCCP should include specific modifications that the
334 manufacturer intends to make over time that generally would otherwise require a new

²⁴ See, e.g., sections 513(i)(1) and 515C of the FD&C Act and 21 CFR 860.7.

²⁵ See FDA’s guidance “[The Least Burdensome Provisions: Concept and Principles.](#)”

²⁶ Sections 513(a)(3)(D)(iv), 513(i)(1)(D)(iii), and 515(c)(5)(D) of the FD&C Act.

²⁷ See FDA’s guidances “[Factors to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications](#)” and “[Benefit-Risk Factors to Consider When Determining Substantial Equivalence in Premarket Notifications \(510\(k\)\) with Different Technological Characteristics.](#)”

²⁸ See sections 515C(a)(1) and 515C(b)(1) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

335 marketing submission.²⁹ A PCCP should not include a list of any/all modifications that a
336 manufacturer may possibly make. To ensure a timely and efficient review, a PCCP
337 should include only a few, specific modifications that can be verified and validated. The
338 modifications included in a PCCP must maintain the device within the device’s intended
339 use,³⁰ and as applicable, must allow the device to remain substantially equivalent to the
340 predicate device.³¹ If a PCCP includes numerous modifications, or modifications that
341 range across various aspects of the device, FDA may not be able to make a determination
342 of reasonable assurance of safety and effectiveness or substantial equivalence for the
343 device and its PCCP.
344

- 345 • **PCCPs harmonize with existing FDA Device Modifications guidances –**
346 Manufacturers can use a PCCP as a way to implement modifications to their devices
347 without needing to submit a new marketing submission for each modification. The
348 Device Modifications guidances help manufacturers determine whether a new marketing
349 submission is required for a modification to their device,³² for example, when the device
350 does not have a PCCP, or when the modification is not consistent with the PCCP. FDA
351 believes this guidance and the Device Modifications guidances support improvement
352 through modifications to devices while continuing to provide a reasonable assurance of
353 device safety and effectiveness.
354

355 **V. Policy for PCCPs**

356 An authorized PCCP specifies planned modifications that, if not included in a PCCP, could
357 otherwise require a new marketing submission pursuant to 21 CFR 807.81(a)(3) and 21 CFR
358 814.39(a), and consistent with the Device Modifications guidances. Because modifications that
359 are specified and implemented in accordance with an authorized PCCP were reviewed and
360 authorized through the marketing submission containing the PCCP, the modifications can be
361 implemented to the device without triggering the need for a new marketing submission.³³
362

363 FDA would consider it to be a deviation from the authorized PCCP in circumstances where the
364 PCCP is not followed or cannot be followed.³⁴ Deviations from the authorized PCCP could
365 significantly affect the safety or effectiveness of the device. This could include, for example,
366 issues related to the Modification Protocol, such as failure to meet pre-specified performance
367 criteria. Device modifications that would not require a marketing submission would not be
368 considered a deviation from an authorized PCCP.³⁵ However, significant modifications made to

²⁹ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

³⁰ See sections 515C(a)(2) and 515C(b)(2) of the FD&C Act.

³¹ Section 515C(b)(2)(B) of the FD&C Act.

³² 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

³³ Sections 515C(a)(1) and 515C(b)(1) of the FD&C Act, 21 CFR 807.81(a)(3), and 21 CFR 814.39(a).

³⁴ FDA would not consider it to be a deviation from the authorized PCCP in situations where a manufacturer chooses not to implement modifications in their authorized PCCP or where a manufacturer chooses to submit a new marketing submission for a device modification in lieu of using their authorized PCCP.

³⁵ See Section V.D. of this guidance for further details on implementing device modifications that may or may not require a new marketing submission in accordance with 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

Contains Nonbinding Recommendations

Draft – Not for Implementation

369 a device that are not specified in, or implemented in accordance with, the authorized PCCP likely
370 require a new marketing submission prior to implementation of the modification.³⁶ Deviations
371 from the authorized PCCP reviewed in the marketing submission would generally cause the
372 device to be adulterated and misbranded under sections 501(f)(1) and 502(o) of the FD&C Act,
373 respectively. The introduction or delivery for introduction into interstate commerce of any food,
374 drug, device, tobacco product, or cosmetic that is adulterated or misbranded is prohibited under
375 section 301(a) of the FD&C Act, and where appropriate, FDA may take legal or regulatory
376 action against violations of prohibited acts, including, without limitation, seizure or injunction.
377

378 **A. Components of a PCCP**

379 A PCCP should consist of a detailed Description of Modifications, a Modification Protocol, and
380 an Impact Assessment (see Section VII. of this guidance) because these components are intended
381 to provide FDA with information that will enable our review of the proposed modifications. The
382 detailed Description of Modifications should outline the specific, planned modifications that may
383 be made to the device. This includes defining the specifications for the characteristics and
384 performance of the planned modifications to the device. The Modification Protocol should
385 describe the verification and validation activities, including pre-defined acceptance criteria, that
386 will support each modification to assure the device remains safe and effective across the intended
387 use populations. The Impact Assessment helps to tie the Description of Modifications to the
388 Modification Protocol in that the Impact Assessment identifies the benefits and risks introduced
389 by the specified, planned modifications and addresses how the verification and validation
390 activities of the Modification Protocol will continue to assure the safety and effectiveness of the
391 device. The detailed Description of Modifications, Modification Protocol, and Impact
392 Assessment are all interrelated components of a PCCP.
393

394 **B. Establishing a PCCP**

395 Premarket authorization for a device with a PCCP must be established through the PMA
396 pathway, 510(k) pathway, or De Novo pathway, as appropriate, as a PCCP must be reviewed and
397 established as part of a marketing authorization^{37,38} for a device prior to a manufacturer
398 implementing any modifications under that PCCP. In general, FDA considers the following
399 submission types to be appropriate³⁹ to establish a PCCP:
400

- 401 • For devices subject to PMA requirements:
 - 402 • Original PMA application
 - 403 • Modular PMA application, where a PCCP comprises a module of review
 - 404 • 180-Day PMA supplement

³⁶ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

³⁷ See sections 513(f)(2) and 515C of the FD&C Act.

³⁸ This includes marketing authorization for a device and PCCP where the device or a derivative thereof has yet to be introduced into interstate commerce, or marketing authorization for a device or a derivative thereof has been introduced into interstate commerce, and for which is being modified to add a PCCP.

³⁹ Submission types for which FDA does not make an affirmative decision (i.e., authorization) would not be appropriate to establish a PCCP.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 405 • 135-Day PMA supplement, where a PCCP comprises certain manufacturing
406 changes only
- 407 • Panel Track PMA supplement
- 408 • Real-Time PMA supplement, where a PCCP comprises minor changes and the
409 manufacturer and FDA agree that the review can be achieved in a real-time
410 setting^{40,41}
- 411 • For devices subject to 510(k) requirements:
 - 412 • Traditional 510(k) submission
 - 413 • Abbreviated 510(k) submission⁴²
- 414 • For devices subject to De Novo requirements:
 - 415 • Original De Novo request

416
417 A PCCP is authorized as part of the device marketing authorization. FDA must reach a
418 determination⁴³ of reasonable assurance of safety and effectiveness or substantial equivalence in
419 review of the device, including each modification specified in the PCCP, for the PCCP to be
420 authorized with the device.

421
422 As manufacturers implement modifications included in an authorized PCCP, FDA expects
423 manufacturers to implement the modifications consistent with their authorized PCCP, i.e., in
424 accordance with their Modification Protocol. A manufacturer must implement any changes in
425 accordance with their quality system.⁴⁴ A manufacturer's quality system is critical for change
426 management processes for a device, especially for devices that include a PCCP, because a PCCP
427 includes modifications that generally would otherwise require a new marketing submission.⁴⁵
428 Compliance with the QSR is essential in order for a manufacturer to implement modifications
429 consistent with their authorized PCCP and failure to do so could potentially present a serious risk
430 to human health.

431
432 Under section 515C(a)(2) of the FD&C Act, FDA may approve a PCCP submitted in a PMA.
433 Under section 515(d)(2)(C) of the FD&C Act, FDA must deny approval of a PMA if FDA finds
434 that the methods used in, or the facilities or controls used for, the manufacture, processing,
435 packing, or installation of such device do not conform to the QSR requirements. Thus, consistent
436 with sections 515C(a)(2) and 515(d)(2)(C) of the FD&C Act, FDA must deny approval of a
437 PCCP submitted in a PMA if FDA finds that the methods used in, or the facilities or controls
438 used for, the manufacture, processing, packing, or installation of the subject device do not
439 conform to the QSR requirements.

440

⁴⁰ Section 737(4)(D) of the FD&C Act defines a real-time supplement as “a supplement to an approved premarket application or premarket report under section 515 that requests a minor change to the device, such as a minor change to the design of the device, software, sterilization, or labeling, and for which the applicant has requested and the agency has granted a meeting or similar forum to jointly review and determine the status of the supplement.”

⁴¹ See FDA's guidance “[Real-Time Premarket Approval Application \(PMA\) Supplements.](#)”

⁴² See FDA's guidance “[The Abbreviated 510\(k\) Program.](#)”

⁴³ See, e.g., sections 513(i)(1) and 515C of the FD&C Act and 21 CFR 860.7.

⁴⁴ 21 CFR Part 820.

⁴⁵ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act, 21 CFR 807.81(a)(3), and 21 CFR 814.39(a).

Contains Nonbinding Recommendations

Draft – Not for Implementation

441 Under section 515C(b)(2) of the FD&C Act, FDA may clear a PCCP submitted in a 510(k).
442 Generally, under section 513(f)(5) of the FD&C Act, FDA may not withhold a determination of
443 the initial classification of a device under section 513(f)(1) of the FD&C Act because of, among
444 other things, a finding that the facility in which the device is manufactured is not in compliance
445 with the QSR. However, also under section 513(f)(5), for devices subject to 510(k), FDA may
446 withhold a substantial equivalence determination if FDA finds that there is a substantial
447 likelihood that the failure to comply with QSR will potentially present a serious risk to human
448 health. Thus, consistent with sections 515C(b)(2) and 513(f)(5) of the FD&C Act, FDA may
449 under certain case-by-case circumstances withhold clearance of a PCCP submitted in a 510(k)
450 based on findings in the regulatory history of the manufacturer that demonstrate failure to
451 comply with QSR.

452
453 For devices subject to 510(k) requirements, the determination of substantial equivalence
454 includes, among other requirements, a comparison between the technological characteristics of
455 the predicate device and the subject device.⁴⁶ In FDA’s determination of substantial equivalence,
456 FDA considers the PCCP to be part of the technological characteristics of the device. For 510(k)
457 submissions, in making a determination of substantial equivalence where the predicate device
458 was authorized with a PCCP, the subject device must be compared to the version of the predicate
459 device cleared or approved prior to changes made under the PCCP.⁴⁷ Once a 510(k) for a device
460 that includes modifications that have been implemented consistent with the authorized PCCP has
461 been cleared in a subsequent marketing submission, such device can now serve as an eligible
462 predicate device. The PCCP can be considered during the 510(k) review process at multiple
463 points in the decision tree to address the critical questions in the 510(k) Decision-Making
464 Flowchart.⁴⁸ In general, FDA anticipates that the PCCP will primarily be reviewed after FDA
465 finds that the intended use of the subject device and the predicate device are the same, to help
466 determine whether the devices have different technological characteristics that do not raise
467 different questions of safety and effectiveness.⁴⁹

468
469 FDA encourages manufacturers to leverage the [Q-Submission Program](#) to obtain FDA feedback
470 on a proposed PCCP for a device and submission type prior to submitting a marketing
471 submission. While manufacturers are encouraged to discuss their plans through a Pre-
472 Submission, FDA does not authorize a PCCP in a Pre-Submission.
473

C. Identifying a PCCP in a Marketing Submission

474
475 The PCCP should be included as a standalone section within the marketing submission, with a
476 title and version number. Additionally, it should be prominently included and discussed in the
477 cover letter and included in the marketing submission’s table of contents as “Predetermined
478 Change Control Plan.” The PCCP should be discussed in the marketing submission as part of the
479 device description, labeling, and relevant sections used for the assessment of reasonable

⁴⁶ See section 513(i) of the FD&C Act.

⁴⁷ See section 515C(c) of the FD&C Act.

⁴⁸ See FDA’s guidance “[The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \[510\(k\)\]](#)” Appendix A, Decision Points 1 through 4.

⁴⁹ See *id.* at Decision Points 5a and 5b.

Contains Nonbinding Recommendations

Draft – Not for Implementation

480 assurance of safety and effectiveness or the substantial equivalence comparison. Any information
481 pertaining to the PCCP content included outside of the PCCP section should include appropriate
482 references to the PCCP section.

483
484 Device labeling must comply with applicable statutes and regulations.⁵⁰ FDA may require that a
485 device with an authorized PCCP include labeling required for safe and effective use of the device
486 as such device changes pursuant to such plan,⁵¹ excluding, as appropriate, trade secret and
487 confidential commercial information. In certain circumstances, for example, when an authorized
488 PCCP is limited to manufacturing changes for a device, it may not be necessary to include
489 information on a device's authorized PCCP in the labeling. However, in the majority of
490 circumstances, information on the device and its authorized PCCP in the labeling is important in
491 order for a user to use the device safely and effectively for the purposes for which it is intended.
492 In particular, information on the device's authorized PCCP may be necessary for a user to
493 understand changes in the device and to continue to use the device safely and effectively across
494 the intended use populations and intended environments as the device changes pursuant to the
495 authorized PCCP.

496
497 As stated above, in the majority of circumstances, FDA recommends that the labeling include a
498 statement that the device has an authorized PCCP. When appropriate, including as modifications
499 are implemented consistent with an authorized PCCP, FDA recommends that the labeling related
500 to the PCCP be updated to include the relevant information listed below for the device and the
501 PCCP so that users may be aware of modifications that have been implemented that impact use
502 of the device:

- 503
- 504 • A description of the implemented modifications, including a summary of current device
505 performance, associated inputs/outputs, validation requirements, and related evidence;
 - 506 • A description of how the modifications were implemented; and
 - 507 • A description of how users will be informed of implemented modifications, including, for
508 example, updated instructions for use or a version history.
- 509

510 When utilizing an authorized PCCP to implement device modifications, the manufacturer should
511 update the labeling for the device as specified in the authorized PCCP.⁵²

512

⁵⁰ 21 CFR Part 801 (Labeling) and 21 CFR Part 809 (In Vitro Diagnostic Products for Human Use). See, e.g., 21 CFR 801.5 (requiring that labeling include adequate directions for use); 21 CFR 801.109(c) (for prescription devices, requiring that labeling include any relevant hazards, contraindications, side effects, and precautions under which practitioners licensed by law to administer the device can use the device safely and for the purpose for which it is intended); and 21 CFR 809.10(b)(6) (for in vitro diagnostic products, requiring labeling accompanying any instruments use or function, installation procedures, performance characteristics and specifications, service and maintenance information, etc.).

⁵¹ See sections 515C(a)(3), 515C(b)(3), and 513(f)(2) of the FD&C Act.

⁵² See Section VII.B.(2)b. for recommendations on update procedures in a Modification Protocol, which should address how labeling will be updated when modifications are implemented, as appropriate.

Contains Nonbinding Recommendations

Draft – Not for Implementation

513 The PCCP should be described in publicly available device summaries, including, for example,
514 the PMA summary of safety and effectiveness document (SSED) and approval order,⁵³ 510(k)
515 summary,^{54,55} or De Novo decision summary.⁵⁶ Details of the PCCP should be included in
516 sufficient detail in the public-facing documents to support transparency to users of the
517 assessment of reasonable assurance of safety and effectiveness or the substantial equivalence
518 comparison for the device and the device’s specifications, excluding, as appropriate, trade secret
519 and confidential commercial information. In addition, FDA recommends public-facing
520 documents include a summary of the following information regarding the PCCP, as appropriate:

- 521
- 522 • Planned modifications;
- 523 • Testing methods;
- 524 • Validation activities and performance requirements to be met in order for modifications
- 525 to be implemented; and
- 526 • Means by which users will be informed of device modifications implemented in
- 527 accordance with the authorized PCCP.
- 528

529 **D. Utilizing an Authorized PCCP to Implement Device** 530 **Modifications**

531 FDA expects manufacturers to follow their authorized PCCP, and manufacturers are required to
532 follow applicable legal requirements regarding the device and its authorized PCCP. In general, a
533 PCCP should be evaluated within the existing risk management framework of the device and
534 must be implemented in accordance with the manufacturer’s quality system.⁵⁷

535
536 Figure 1 depicts the process for implementing a modification to a device with an authorized
537 PCCP. Manufacturers should first consider whether the particular modification is or is not
538 consistent with the authorized PCCP; FDA considers a modification to be consistent with the

⁵³ In accordance with 21 CFR 814.9(e), “FDA will make available to the public ... a detailed summary of information submitted to FDA respecting the safety and effectiveness of the device that is the subject of the PMA and that is the basis for the order.”

⁵⁴ In accordance with 21 CFR 807.92, “a 510(k) summary shall be in sufficient detail to provide an understanding of the basis for a determination of substantial equivalence.” This includes, but is not limited to, a description of the device, and for those 510(k) submissions in which a determination of substantial equivalence is also based on an assessment of performance data, non-clinical tests, and clinical tests.

⁵⁵ If a sponsor chooses to submit a 510(k) Statement rather than a 510(k) Summary, the sponsor should provide to requestors all PCCP information in the 510(k) that supports transparency to users of FDA’s determination of substantial equivalence for the device and its specifications, as such information constitutes safety and effectiveness information. See 21 CFR 807.93 for requirements on the content and format of a 510(k) Statement.

⁵⁶ The De Novo decision summary is intended to present an objective and balanced summary of the scientific evidence that served as the basis for the FDA’s decision to grant a De Novo request; see FDA’s website on [De Novo Classification Request](#).

⁵⁷ Manufacturers are required to comply with the QSR (21 CFR Part 820). The device and PCCP must be implemented consistent with 21 CFR Part 820, including, but not limited to design controls (21 CFR 820.30), nonconforming products (21 CFR 820.90), and corrective and preventative action (21 CFR 820.100). The QSR also includes requirements to review and approve modifications to device design and production (21 CFR 820.30 and 820.70), and requirements to document changes and approvals in the device master record (21 CFR 820.181).

Contains Nonbinding Recommendations

Draft – Not for Implementation

539 authorized PCCP when the modification has been specified in the Description of Modifications
540 included in the PCCP *and* has been implemented in accordance with the Modification Protocol.
541 If the particular modification is consistent with the authorized PCCP, a new marketing
542 submission is not necessary; the modification can be implemented in accordance with the
543 Modification Protocol, and the manufacturer should document that modification and the analysis
544 in accordance with the manufacturer’s quality system.
545

546 If the particular modification is not consistent with the authorized PCCP – including if the
547 modification is not included in the authorized PCCP or if the modification is included in the
548 authorized PCCP but is not implemented in accordance with the methods and specifications
549 described in the Modification Protocol – the manufacturer should then proceed to evaluate the
550 particular modification as described below. As part of the review of the particular modification,
551 manufacturers should consider how implementation of the particular modification may affect the
552 PCCP for the device by reviewing the Impact Assessment of the PCCP to determine if the
553 particular modification introduces or significantly modifies risk mitigations for the PCCP.
554

- 555 • If the modification is not included in the authorized PCCP, the manufacturer should
556 proceed based on their evaluation of the particular modification in accordance with
557 applicable FDA statutory and regulatory requirements and after consulting the Device
558 Modifications guidances.
- 559 • If the modification is included in the authorized PCCP but is not implemented in
560 accordance with the methods and specifications described in the Modification Protocol,
561 in most cases, because modifications included in a PCCP are those that would generally
562 otherwise require a new marketing submission, it is likely that a new marketing
563 submission will be required before the manufacturer can implement the change.⁵⁸
564

565 As described in the introduction of Section V. of this guidance, FDA would consider it to be a
566 deviation from the authorized PCCP in circumstances where the PCCP is not followed or cannot
567 be followed.⁵⁹ Deviations from the authorized PCCP could significantly affect the safety or
568 effectiveness of the device. Significant modifications made to a device that are not specified in,
569 or implemented in accordance with, the authorized PCCP likely require a new marketing
570 submission.⁶⁰ If a manufacturer believes that the deviation from their authorized PCCP is not
571 significant, we strongly encourage discussion with the appropriate FDA review division. In
572 general, manufacturers may contact the appropriate FDA review division for a discussion about a
573 proposed modification and whether it may be considered consistent with the authorized PCCP.
574

575 If, after review of applicable FDA statutory and regulatory requirements a new marketing
576 submission is required,⁶¹ the appropriate marketing submission could request authorization for:
577

⁵⁸ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

⁵⁹ FDA would not consider it to be a deviation from the authorized PCCP in situations where a manufacturer chooses not to implement modifications in their authorized PCCP or where a manufacturer chooses to submit a new marketing submission for a device modification in lieu of using their authorized PCCP.

⁶⁰ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

⁶¹ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 578 • A modification to the authorized PCCP⁶² (see Section V.E. of this guidance); and/or
579 • The modified device (i.e., a device modification not implemented through a PCCP).
580

581 In each of these cases, the marketing submission for the modification must include the
582 appropriate marketing submission requirements⁶³ for the device. If the manufacturer requests
583 authorization for a modification to the authorized PCCP, the manufacturer must also submit the
584 proposed, modified PCCP for the device.⁶⁴ If the manufacturer requests authorization for the
585 modified device, the manufacturer must also submit the proposed PCCP for the modified
586 device.⁶⁵ In both scenarios, the manufacturer must obtain FDA authorization for the device and
587 proposed PCCP before implementing the PCCP.⁶⁶
588

⁶² A change to the authorized PCCP could include a change in Description of Modifications, the Modification Protocol, and/or the Impact Assessment, as appropriate.

⁶³ See, e.g., 21 CFR 807.87, 21 CFR 860.220, or 21 CFR 814.20. In general, manufacturers may provide references in the marketing submission to prior marketing submissions for content that remains unchanged, as appropriate.


⁶⁴ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act.

⁶⁵ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act.

⁶⁶ Section 515C of the FD&C Act, 21 CFR 807.81(a)(3), and 21 CFR 814.39(a).

Contains Nonbinding Recommendations

Draft – Not for Implementation



¹ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

² See the FDA guidances "Modifications to Devices Subject to Premarket Approval (PMA) – The PMA Supplement Decision-Making Process," "Deciding When to Submit a 510(k) for a Change to an Existing Device," or "Deciding When to Submit a 510(k) for a Software Change to an Existing Device."

³ For the modified device to have a PCCP, a PCCP should be submitted with the marketing submission so that the device and PCCP can be authorized together.

This flowchart is not intended to be used as a 'stand-alone' document and should only be considered in conjunction with the accompanying text in this guidance.

589
590
591
592

Figure 1. Implementing a Modification to a Device with an Authorized PCCP

593

E. Modifying a Previously Authorized PCCP

594 FDA believes that modifications to an authorized PCCP will generally constitute changes to the
595 device that would otherwise require a new marketing submission.⁶⁷ In other words, FDA
596 anticipates that modifications to a PCCP will need to be reviewed and established as part of the
597 marketing submission for the modified device because a modification to the PCCP will generally
598 significantly affect the safety or effectiveness of the device.⁶⁸

⁶⁷ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act, 21 CFR 807.81(a)(3), and 21 CFR 814.39(a).

⁶⁸ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act, 21 CFR 807.81(a)(3), and 21 CFR 814.39(a).

Contains Nonbinding Recommendations

Draft – Not for Implementation

599
600 For a manufacturer who would like to modify their PCCP for a previously authorized device
601 with a PCCP,⁶⁹ the marketing submission must include the appropriate marketing submission
602 requirements⁷⁰ and the proposed, modified PCCP for the device.⁷¹ We recommend that
603 manufacturers provide a summary of the changes to the authorized PCCP, and where practicable,
604 a tracked changes version compared to the authorized PCCP. In general, FDA considers the
605 PMA supplement and 510(k) submission types included in Section V.B. of this guidance to be
606 appropriate to modify a PCCP. In addition to those submission types, for devices subject to
607 510(k) requirements, a special 510(k) submission may be appropriate to modify a PCCP where
608 the modifications to a PCCP comprise changes to the manufacturer’s own device and PCCP and
609 where well-established methods are available to evaluate the change to the PCCP.⁷²

610
611 FDA intends to focus its review on the aspects of the device that are most significantly
612 modified.⁷³ For example, if the device has been relatively unchanged since FDA’s prior
613 authorization and a modified PCCP is proposed, FDA would focus its review on the modified
614 PCCP. Manufacturers may discuss proposed changes to the PCCP with the appropriate FDA
615 review division through the [Q-Submission Program](#).

617 **F. Version Control and Maintenance of a PCCP for a Device**

618 At this time, as manufacturers gain experience developing and implementing PCCPs, FDA
619 anticipates that review of a marketing submission that includes a PCCP will be interactive. It is
620 possible that a PCCP submitted as part of a marketing submission may need revisions before
621 FDA can make a determination⁷⁴ of reasonable assurance of safety and effectiveness or
622 substantial equivalence in review of the device, including each modification in the PCCP, for the
623 PCCP to be authorized with the device. FDA should work with the manufacturer to revise the
624 PCCP, and will do so using FDA’s existing processes to request additional information through
625 interactive review or through a deficiency letter.⁷⁵ If deficiencies with the PCCP remain
626 unresolved, FDA may authorize the device upon withdrawal of the PCCP.

627
628 As described in Section V.C. of this guidance, a copy of the PCCP with a title and version
629 number should be included in the marketing submission for the device. If the PCCP is revised
630 during review, such as in response to deficiencies, a final, revised version of the PCCP should be
631 submitted as a clean copy to FDA and should include a title and current version number for the
632 PCCP. FDA authorizes the PCCP as part of the marketing authorization of a device. To that end,

⁶⁹ E.g., through a PMA supplement or a traditional 510(k) for a device that has already been authorized.

⁷⁰ See, e.g., 21 CFR 807.87, 21 CFR 860.220, or 21 CFR 814.20. In cases where the modified PCCP is the reason for the marketing submission, and in general, manufacturers may provide references in the marketing submission to prior marketing submissions for content that remains unchanged, as appropriate.

⁷¹ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act.

⁷² See FDA’s guidance “[The Special 510\(k\) Program](#).”

⁷³ Note that “focus of the review” is not intended to imply a review of the PCCP *only*; rather, the focus on the PCCP is as a significant change to the device that could affect the safety or effectiveness of the device.

⁷⁴ See, e.g., sections 513(i)(1) and 515C of the FD&C Act and 21 CFR 860.7.

⁷⁵ See FDA’s guidance, “[Developing and Responding to Deficiencies in Accordance with the Least Burdensome Provisions](#).”

Contains Nonbinding Recommendations

Draft – Not for Implementation

633 the PCCP will be referenced in the device’s letter of authorization, including the PCCP title and
634 version number. A manufacturer should only have one version of an authorized PCCP for their
635 device. However, a PCCP can evolve over time through future marketing submissions where a
636 new version of the PCCP can be authorized. As such, there should only be one version of the
637 PCCP under review with a device at any given time to help with version control of the PCCP for
638 manufacturers and FDA.

639
640 Over time, as a manufacturer implements their authorized PCCP for their device, they may wish
641 to make modifications to the device that are not included in their authorized PCCP and that may
642 require a new marketing submission.⁷⁶ If prior modifications have been implemented consistent
643 with the authorized PCCP for the device, the new marketing submission for the device should
644 include a summary of those modifications that have been implemented. This may include
645 information in the device description, labeling, and other relevant sections of the marketing
646 submission so that FDA can understand the current device characteristics and performance. In
647 the context of premarket authorization, FDA does not intend to re-review the adequacy of
648 modifications implemented consistent with an authorized PCCP; however, FDA needs to have an
649 understanding of the current device characteristics and performance in order to make a
650 determination⁷⁷ of reasonable assurance of safety and effectiveness or substantial equivalence.
651 Additionally, if the manufacturer would also like to modify their previously authorized PCCP,
652 the marketing submission must include the appropriate marketing submission requirements⁷⁸ and
653 the proposed, modified PCCP for the device (see Section V.E. of this guidance).⁷⁹

654
655 For devices subject to PMA requirements, submission of periodic reports, at intervals of one year
656 (unless otherwise specified) from the date of approval of the original PMA are required under 21
657 CFR 814.84. The Annual Report must include, separately for each model number (if applicable),
658 the number of devices sold and distributed during the reporting period, including those
659 distributed to distributors.⁸⁰ FDA recommends that the Annual Report include a separate section
660 that describes any changes implemented through an approved PCCP. Additionally, for devices
661 subject to PMA requirements, when a modification implemented consistent with the authorized
662 PCCP necessitates an update to the labeling, manufacturers should provide a summary of the
663 updated labeling in the annual report.⁸¹

664
665 For devices subject to 510(k) requirements, a manufacturer must compare their subject device to
666 the version of the predicate device cleared or approved prior to changes made under the PCCP.⁸²
667 However, once a 510(k) for a device that includes modifications that have been implemented

⁷⁶ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

⁷⁷ See, e.g., sections 513(i)(1) and 515C of the FD&C Act and 21 CFR 860.7.

⁷⁸ See, e.g., 21 CFR 807.87, 21 CFR 860.220, or 21 CFR 814.20. In general, manufacturers may provide references in the marketing submission to prior marketing submissions for content that remains unchanged, as appropriate.

⁷⁹ Sections 515C(a)(2) and 515C(b)(2) of the FD&C Act.

⁸⁰ See 21 CFR 814.82(a)(9) (noting that “[s]uch other requirements as FDA determined are necessary to provide reasonable assurance, or continued reasonable assurance, of the safety and effectiveness of the device” may be included as post-approval requirements for a PMA-approved device).

⁸¹ See FDA’s guidance “[Annual Reports for Approved Premarket Approval Applications \(PMA\)](#).”

⁸² Section 515C(c) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation

668 consistent with the authorized PCCP has been cleared in a subsequent marketing submission,
669 such device can now serve as an eligible predicate device.
670

671 **VI. Types of Modifications**

672 Modifications that are appropriate for inclusion in a PCCP include those that are intended to
673 maintain or improve the safety or effectiveness of the device. Modifications should also be
674 specific, and should be able to be verified and validated.
675

676 Modifications included in a PCCP must maintain the device within the device’s intended use,⁸³
677 and as applicable, must allow the device to remain substantially equivalent to the predicate
678 device.⁸⁴ In general, FDA believes that modifications included in a PCCP should also maintain
679 the device within the device’s indications for use.⁸⁵ As with modifications to the intended use,
680 FDA believes that most modifications to the indications for use included in a PCCP would be
681 difficult for FDA to assess prospectively to determine whether the device would remain safe and
682 effective.⁸⁶ However, there may be certain modifications to the indications for use that may be
683 appropriate for inclusion in a PCCP, and are detailed further in the subsections below. FDA
684 highly encourages manufacturers discuss modifications to the indications for use that may be
685 included in a proposed PCCP with the appropriate FDA review division through the [Q-
686 Submission Program](#).
687

688 In the subsections below, FDA provides recommendations for how manufacturers should
689 determine whether a modification may be appropriate for inclusion in a PCCP. We also provide
690 some high-level examples of modifications that generally may be or are not appropriate for
691 inclusion in a PCCP. However, the high-level examples are not exhaustive, and are not intended
692 to cover all possible details, risks, or considerations. Ultimately, decisions about the types of
693 modifications to be included in a PCCP are generally fact-specific for each device.
694

695 Recognizing there is a spectrum of risk for devices, for purposes of reviewing a PCCP, FDA
696 intends to, among other considerations, take into account the Guiding Principles recommended in
697 this guidance (Section IV.), the benefit-risk profile of the specific device that is the subject of the
698 PCCP, the specific modifications being proposed, and FDA’s experience applying this policy
699 across different device types. As such, certain modifications that may be appropriate for
700 inclusion in a PCCP for one device may not be appropriate for inclusion in a PCCP for another
701 device. Some modifications may not be appropriate for inclusion within a PCCP for any device.
702

⁸³ See sections 515C(a)(2) and 515C(b)(2) of the FD&C Act.

⁸⁴ Section 515C(b)(2)(B) of the FD&C Act.

⁸⁵ FDA has a long-standing policy of applying the definition of indications for use in the PMA regulation at 21 CFR 814.20(b)(3)(i) in the same way in the 510(k) context. See the FDA guidance “[The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket Notifications \[510\(k\)\]](#).”

⁸⁶ Sections 515C(a)(2) and 515C(b)(2)(A) of the FD&C Act.

703 **A. Determining Whether a Modification may be Appropriate**
704 **for Inclusion in a PCCP in a 510(k) or De Novo**
705 **Submission**

706 Pursuant to 21 CFR 807.81(a)(3), and as further described in FDA’s guidances on “[Deciding](#)
707 [When to Submit a 510\(k\) for a Change to an Existing Device](#)” and “[Deciding When to Submit a](#)
708 [510\(k\) for a Software Change to an Existing Device](#)” (referred to hereafter as “510(k)
709 Modifications guidances”), “significant changes or modifications that require a [510(k)]” include
710 “[a] change or modification that could significantly affect the safety or effectiveness of the
711 device, e.g., a significant change or modification in design, material, chemical composition,
712 energy source, or manufacturing process” or “[a] major change or modification in the intended
713 use of the device.” Consistent with section 515C(b)(2) of the FD&C Act, FDA “may clear a
714 [PCCP] submitted in a [510(k)] that describes planned changes that may be made to the device
715 (and that would otherwise require a new [510(k)]), if—(A) the device remains safe and effective
716 without any such change; and (B) the device would remain substantially equivalent to the
717 predicate.”

718
719 Under section 515C(b)(2)(B) of the FD&C Act, FDA may clear a PCCP for planned changes that
720 would otherwise require a new 510(k) if, among other things, the device remains substantially
721 equivalent to the predicate. Under section 513(i)(1)(A) of the FD&C Act, substantial equivalence
722 means, in part, that the device has the same intended use as the predicate device. Therefore, the
723 modifications included in a PCCP must maintain the device within the device’s intended use.⁸⁷
724 However, a significant change or modification in design, material, chemical composition, energy
725 source, or manufacturing process may be appropriate for inclusion in a PCCP. As described in
726 the 510(k) Modifications guidances, FDA recommends when assessing such types of
727 modifications that a manufacturer should first conduct a risk-based assessment to determine
728 whether the modification would require a new marketing submission,⁸⁸ and therefore whether
729 that modification may be appropriate for inclusion in a PCCP. FDA recommends that
730 manufacturers use the results of a risk-based assessment to help categorize modifications by
731 those that could introduce new risks or those that could significantly modify existing risks.

732
733 Accordingly, modifications that could introduce a new risk are those that could introduce a new
734 hazard or hazardous situations that did not exist for the original device and for which the pre-
735 mitigation risk level associated with the new risk is not considered to be acceptable.
736 Modifications that could introduce new risks are generally not appropriate for inclusion in a
737 PCCP because the risks of implementing the modification are likely not adequately mitigated by
738 the existing risk management framework of the device and the manufacturer’s quality system.

739
740 Modifications that could significantly modify an existing risk are those that could change the risk
741 score, risk acceptability category, or duration of risk. Modifications that could significantly
742 modify existing risks generally may be appropriate for inclusion in a PCCP when the risks of
743 implementing the modification are adequately mitigated by the existing risk management

⁸⁷ Section 515C(b)(2) of the FD&C Act.


⁸⁸ 21 CFR 807.81(a)(3) and 21 CFR 814.39(a).

Contains Nonbinding Recommendations

Draft – Not for Implementation

744 framework of the device and the manufacturer’s quality system. For additional information on
745 whether a modification could introduce a new risk or significantly change an existing risk, see
746 the 510(k) Modifications guidances.

747
748 To harmonize with our policy for device modifications for devices subject to 510(k) or De Novo
749 requirements, FDA recommends using the following process (see Figure 2) to determine whether
750 a modification may be appropriate for inclusion in a PCCP. We recommend that the process
751 depicted in Figure 2 be considered in conjunction with the recommendations in this guidance, as
752 there are many factors to consider in concert to determine whether a modification may be
753 appropriate for inclusion in a PCCP.
754



This flowchart is not intended to be used as a ‘stand-alone’ document and should only be considered in conjunction with the accompanying text in this guidance.

755
756
757
758
759

Figure 2. Determining Whether a Modification may be Appropriate for Inclusion in a PCCP for a 510(k) or De Novo Device

760 Please see below for lists of certain high-level modifications that generally may be appropriate or
761 are not appropriate for inclusion in a PCCP. To avoid misinterpretation, we recommend that the
762 lists be considered together to determine the appropriateness of including a modification in a
763 PCCP. For example, a change in materials may be appropriate to include in a PCCP, however, if
764 such a change may need new clinical data, it would not be appropriate to include in a PCCP.

765

Modifications that generally may be appropriate for inclusion in a PCCP:

766
767

- Certain changes in device design, including dimensions, performance specifications, wireless communication, components or accessories, or the patient/user interface

768
769

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 770 • Change in sterilization, packaging, transport, or expiration dating using well-established
771 methods
- 772 • Certain changes in materials/components (e.g., different raw materials, reagents, or
773 hardware components)
- 774 • Certain changes in software related to device compatibility and/or interoperability (e.g.,
775 changes to support device use on additional operating system(s), new data vendors and/or
776 sources, or compatibility with additional devices)
- 777 • Certain changes in software consistent with the intended use to improve device
778 performance
- 779 • Certain changes to the labeling to describe a specific subset of a patient population within
780 the originally indicated patient population that the device is intended for use in
781 diagnosing, treating, preventing, curing, or mitigating
- 782 • Certain changes in the labeling and/or the indications for use to specify use of the device
783 with an additional device, component, or human genetic variant
- 784 • Certain changes in the indications for use regarding use in the home setting
785

Modifications that are generally not appropriate for inclusion in a PCCP:

- 786 • Change to device control mechanism, operating principle, or energy type
- 787 • Change in device design that could affect the intended use of the device
- 788 • Change from a device labeled for single use only to a device labeled as reusable
- 789 • Change to or removal of contraindications
- 790 • Change from prescription to over-the-counter use
- 791 • Changes from “general to specific”⁸⁹
- 792 • Change in the labeling and/or the indications for use to include a new patient population
- 793 • Changes that may need new clinical data⁹⁰
- 794 • Change to address a recall or safety issue
- 795 • Change to a device constituent part that impacts the biologic or drug constituent part
796
797
798

B. Determining Whether a Modification may be Appropriate for Inclusion in a PCCP in a PMA Application or Supplement

802 Pursuant to section 515(d)(5)(A)(i) of the FD&C Act and 21 CFR 814.39(a), and as further
803 described in FDA’s guidance on “[Modifications to Devices Subject to Premarket Approval](#)
804 [\(PMA\) - The PMA Supplement Decision-Making Process](#)” (referred to hereafter as the “PMA
805 Modifications guidance”), “[a]fter FDA’s approval of a PMA, an applicant shall submit a PMA
806 supplement for review and approval by FDA before making a change affecting the safety or

⁸⁹ A change from “general to specific”, i.e., a change from general to a specific indications for use, is any proposed increase in the level of specificity of the indication for use of a medical device. For additional information, see FDA’s guidance “[General/Specific Intended Use](#).”

⁹⁰ Certain changes that may need new clinical data, such as method comparison data for IVDs, may be appropriate for inclusion in a PCCP.

Contains Nonbinding Recommendations

Draft – Not for Implementation

807 effectiveness of the device for which the applicant has an approved PMA” unless an exception
808 applies. Consistent with section 515C(a)(2) of the FD&C Act, FDA “may approve a [PCCP]
809 submitted in an application, including a supplemental application, under section 515 that
810 describes planned changes that may be made to the device (and that would otherwise require a
811 supplemental application under section 515), if the device remains safe and effective without any
812 change.”

813
814 Under section 515C(a)(2) of the FD&C Act, FDA may approve a PCCP for planned changes that
815 would otherwise require a supplemental application under section 515 of the FD&C Act.
816 Therefore, the modifications included in a PCCP must maintain the device within the device’s
817 intended use.⁹¹ Other modifications that could affect the safety or effectiveness of the device
818 may be appropriate for inclusion in a PCCP.

819
820 Minor changes and manufacturing changes are two types of modifications that could affect the
821 safety or effectiveness of the device. A modification is considered a minor change if it is a minor
822 change to the design of the device, software, sterilization, or labeling, typically otherwise
823 reviewed under a real-time supplement.⁹² A modification is a manufacturing change if it is a
824 modification to the manufacturing procedures or methods of manufacture affecting the safety or
825 effectiveness of the device, typically otherwise reviewed under a 30-day notice.⁹³

826
827 Accordingly, modifications that are minor changes or manufacturing changes generally may be
828 appropriate for inclusion in a PCCP when the risks of implementing the modification are
829 adequately mitigated by the existing risk management framework of the device and the
830 manufacturer’s quality system. Conversely, modifications that are not minor changes or
831 manufacturing changes are generally not appropriate for inclusion in a PCCP because the risks of
832 implementing the modification are likely not adequately mitigated by the existing risk
833 management framework of the device and the manufacturer’s quality system.

834
835 To harmonize with our policy for device modifications for devices subject to PMA requirements,
836 FDA recommends using the following process (see Figure 3) to determine whether a
837 modification may be appropriate for inclusion in a PCCP. We recommend that the process
838 depicted in Figure 3 be considered in conjunction with the recommendations in this guidance, as
839 there are many factors to consider in concert to determine whether a modification may be
840 appropriate for inclusion in a PCCP.

841


⁹¹ See section 515C(a)(2) of the FD&C Act.

⁹² See section 737(4)(D) of the FD&C Act.

⁹³ See section 737(5) of the FD&C Act.

Contains Nonbinding Recommendations

Draft – Not for Implementation



This flowchart is not intended to be used as a 'stand-alone' document and should only be considered in conjunction with the accompanying text in this guidance.

842
843

844 **Figure 3. Determining Whether a Modification may be Appropriate for Inclusion in a** 845 **PCCP for a PMA Device**

846

847 Please see below for lists of certain high-level modifications that generally may be appropriate or
848 are not appropriate for inclusion in a PCCP. To avoid misinterpretation, we recommend that the
849 lists be considered together to determine the appropriateness of including a modification in a
850 PCCP. For example, a minor change in the instructions for use in the labeling may be appropriate
851 to include in a PCCP, however, if such a change addresses a safety issue, it would not be
852 appropriate to include in a PCCP.

853

854 **Modifications that generally may be appropriate for inclusion in a PCCP:**

855

- 856 • Minor change in device design, including dimensions, performance specifications,
857 wireless communication, components or accessories, or the patient/user interface
- 858 • Minor change in sterilization, packaging, transport, or expiration dating
- 859 • Minor change in a material/component that has similar technical specifications to those
860 for the authorized device (e.g., different source or supplier for raw materials, reagents, or
861 hardware components)
- 862 • Minor change in software related to device compatibility and/or interoperability (e.g.,
863 changes to support device use on upgraded operating system(s) or new data vendors
864 and/or sources)
- 865 • Minor change in software consistent with the intended use to improve device
866 performance

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 867 • Minor change to the labeling to describe a specific subset of a patient population within
868 the originally indicated patient population that the device is intended for use in
869 diagnosing, treating, preventing, curing, or mitigating
- 870 • Minor change to the labeling, including instructions for use, warnings, precautions, or
871 other labeling that does not affect the indications for use or contraindications
- 872 • Minor change in the labeling and/or the indications for use to specify use of the device
873 with an additional device, component, or human genetic variant
- 874 • Certain changes in manufacturing procedures:
 - 875 • Change to sterilization, joining, or cleaning procedures
 - 876 • Change to automate existing processes
 - 877 • Change to environmental conditions of manufacturing, storage, or distribution
878 facilities (e.g., addition/relocation of a room for manufacturing purposes to the
879 existing manufacturing space)
- 880 • Certain changes in methods of manufacture:
 - 881 • Change in manufacturing materials (e.g., new/different machine lubricants)
 - 882 • Change in manufacturing software
 - 883 • Change in supplier for manufacturing components where specifications are
884 unchanged (e.g., addition of a raw material supplier for materials critical to the
885 performance of the device)
 - 886

Modifications that are generally not appropriate for inclusion in a PCCP:

- 887 • Significant change to components, materials, design, specifications, software, or color
888 additives, such as:
 - 889 • Change to device control mechanism, operating principle, or energy type
 - 890 • Change in device design that could affect the intended use of the device
- 891 • Significant change in labeling, such as:
 - 892 • Change from a device labeled for single use only to a device labeled as reusable
 - 893 • Changes from “general to specific”⁹⁴
 - 894 • Change to or removal of contraindications
- 895 • Change in the labeling and/or the indications for use to include a new patient population
- 896 • Changes that may need new clinical data⁹⁵
- 897 • Change to address a recall or safety issue
- 898 • Change to a device constituent part that impacts the biologic or drug constituent part
- 899 • Change to add, expand, or move the manufacturing or sterilization site of a finished
900 device
- 901
- 902
- 903

VII. Recommended Content for a PCCP

A. Description of Modifications

⁹⁴ See footnote 89.

⁹⁵ See footnote 90.

Contains Nonbinding Recommendations

Draft – Not for Implementation

906 A description of each planned modification to a device should be included in the Description of
907 Modifications section of a PCCP. The detailed description should describe specific changes to
908 the device characteristics and performance resulting from implementation of the modifications.
909 To ensure an efficient review, FDA recommends that a PCCP include only a limited number of
910 modifications that are specific, and that can be verified and validated.
911

912 (1) Goals of the Description of Modifications

913 A dedicated Description of Modifications section in a PCCP should identify the specific, planned
914 modifications to the device that the manufacturer intends to implement. The Description of
915 Modifications should include the specifications for the characteristics and performance of the
916 device that, following the agreed upon verification and validation described in the Modification
917 Protocol, can be implemented without a new marketing submission.
918

919 (2) Content of the Description of Modifications

920 To achieve these goals, FDA recommends that the Description of Modifications enumerate the
921 list of individual proposed device modifications discussed in the PCCP, as well as the specific
922 rationale for the change to each part of the device that is planned to be modified. In some
923 situations, a Description of Modifications may consist of multiple modifications. It may be
924 helpful to reference the labeling sections that are anticipated to be impacted for each
925 modification in the Description of Modifications section (such labeling changes should be
926 included in the Modification Protocol, as described in Section VII.B.(2)b. of this guidance).
927

928 FDA recommends that a PCCP include modifications that are specific, and that can be verified
929 and validated. Modifications should also be presented at a level of detail that permits
930 understanding of the specific modifications that will be made to the device. Each modification
931 should be linked to a specific performance evaluation activity within the Modification Protocol
932 (for an example, see Table 1 in Section VII.C. of this guidance).
933

934 **B. Modification Protocol**

935 The Modification Protocol should include the documentation describing the methods that will be
936 followed when developing, validating, and implementing modifications delineated in the
937 Description of Modifications section of the PCCP. The Modification Protocol should also
938 include the verification and validation activities, including pre-defined acceptance criteria, for
939 those modifications, and a step-by-step delineation of how those modifications will be
940 implemented while assuring the device remains safe and effective.
941

942 Documentation of modifications verified and validated per the Modification Protocol must be
943 compliant with the QSR, including that the manufacturer must document the change in
944 accordance with the manufacturer's quality system.⁹⁶ The QSR requires manufacturers of
945 finished medical devices to review and approve modifications to device design and production

⁹⁶ 21 CFR Part 820.

Contains Nonbinding Recommendations

Draft – Not for Implementation

946 (21 CFR 820.30(i) and 820.70(b)) and document changes and approvals in the device master
947 record (21 CFR 820.181).
948

949 **(1) Goals of the Modification Protocol**

950 Whereas the Description of Modifications outlines the planned modifications to a device, the
951 Modification Protocol should describe the methods that will be followed when developing,
952 validating, and implementing those modifications to ensure the device remains safe and
953 effective. The methods described in the Modification Protocol should be consistent with and
954 support the modifications outlined in the Description of Modifications.
955

956 The goals of the Modification Protocol are to:
957

- 958 • Identify the appropriate and applicable data, test methods, analysis methods, and
959 specified acceptance criteria used to develop, validate, and implement all proposed
960 modifications;
- 961 • Identify the update process, and as appropriate, the plans for communication and/or
962 training for users for implemented modifications;
- 963 • Ensure that the information that would otherwise be generated and submitted to FDA
964 (i.e., if the modifications were implemented on a device that did not have an
965 authorized PCCP) will be generated by the manufacturer for each modification and
966 maintained consistent with recordkeeping requirements and in accordance with the
967 manufacturer's quality system;⁹⁷
- 968 • Ensure that anticipated risks and risk mitigation strategies have been identified and
969 included in the Impact Assessment; and
- 970 • Be least burdensome⁹⁸ for the manufacturer to develop and for FDA to review. This
971 includes being traceable (so that modifications in the Description of Modifications
972 can be traced to verification and validation activities in the Modification Protocol)
973 and specific (so that the PCCP is sufficiently comprehensive to support the proposed
974 modifications).
975

976 Manufacturers should follow their risk management processes to develop a Modification
977 Protocol that considers each modification. In some cases, the same methods described in the
978 Modification Protocol may support all modifications in a PCCP for a device. In other cases, the
979 same methods described in the Modification Protocol may not be adequate for every
980 modification in a PCCP. For each planned modification provided in the Description of
981 Modifications, FDA recommends that the information outlined in this section be addressed in a
982 Modification Protocol. The Modification Protocol should include a description of how its
983 proposed methods are similar to or different from methods used elsewhere in the marketing
984 submission. For example, if the validation methods in the Modification Protocol represent a
985 subset of the original testing for the device, or if the acceptance criteria for the validation are
986 different, manufacturers should describe these differences and provide a justification. The

⁹⁷ 21 CFR Part 820.

⁹⁸ See FDA's guidance "[The Least Burdensome Provisions: Concept and Principles.](#)"

Contains Nonbinding Recommendations

Draft – Not for Implementation

987 justification for a different methodology may include references to other marketing submissions
988 where the methodology was used for similar modifications.
989

990 **(2) Content of the Modification Protocol Section**

991 To achieve these goals, FDA recommends that a Modification Protocol include information
992 regarding the manufacturer’s performance evaluation methods, and also, when appropriate,
993 update procedures. In FDA’s experience, this is generally the type of information that will enable
994 FDA to evaluate the PCCP. For a particular marketing submission, additional information in a
995 Modification Protocol may need to be included to assist in FDA’s determination of substantial
996 equivalence or reasonable assurance of safety and effectiveness of the device and PCCP.
997

998 **a. Performance Evaluation Methods**

999 Performance evaluation of the device is important to ensure that specified acceptance criteria for
1000 all proposed modifications will continue to be met for the device’s specifications. FDA may
1001 require that performance requirements for changes made under the plan be provided in a
1002 PCCP.⁹⁹ Performance evaluation methods should include the plans to verify and validate that the
1003 modified device will meet the specifications identified as part of a specific modification, in
1004 addition to maintaining the specifications that are not part of the modification, but may be
1005 impacted by the modification. Performance evaluation should include, as applicable, the plans
1006 for verification and validation of the entire device following the implementation of each
1007 individual modification and in aggregate for the planned modifications. In general, depending on
1008 the proposed modifications, this information may be similar to the performance evaluation
1009 methods used to support the original marketing submission for the device. The content of this
1010 section in a Modification Protocol should provide details on the study design, performance
1011 metrics, pre-defined acceptance criteria, and statistical tests for each planned modification. More
1012 comprehensive testing can potentially support a broader set of proposed modifications.
1013

1014 To determine the information that manufacturers should provide in a Modification Protocol,
1015 FDA recommends considering the specific, planned modifications to your device and reviewing
1016 FDA guidances, including, but not limited to:

- 1017
- 1018 • Device-specific guidance¹⁰⁰ that may be applicable to your device;
 - 1019 • Performance testing guidances, such as those on non-clinical bench performance testing
1020 or analytical studies¹⁰¹ and non-clinical animal performance testing,^{102,103}; and

⁹⁹ See sections 515C(a)(3), 515C(b)(3), and 513(f)(2) of the FD&C Act.

¹⁰⁰ See [FDA’s guidance search database](#).

¹⁰¹ See, e.g., FDA’s guidance on “[Recommended Content and Format of Non-Clinical Bench Performance Testing Information in Premarket Submissions](#).”

¹⁰² FDA supports the principles of the “3Rs,” to replace, reduce, and/or refine animal use in testing, when feasible. We encourage manufacturers to consult with FDA if they wish to use a non-animal testing method that they believe is suitable, adequate, validated, and feasible. We will consider if a proposed alternative method could be assessed for equivalency to an animal test method.

¹⁰³ See, e.g., FDA’s guidance on “[General Considerations for Animal Studies Intended to Evaluate Medical Devices](#).”

C. Traceability Between the Description of Modifications Section and the Modification Protocol Section

The PCCP should clearly delineate which parts of the Modification Protocol are applicable to each modification within the Description of Modifications. For a PCCP with multiple modifications, this may be accomplished through a traceability table; a sample traceability table is provided below in Table 1. This sample traceability table provides an example of how a manufacturer can depict the traceability between the Description of Modifications and Modification Protocol, as well as how to provide clear references to where within the PCCP this information is located in a marketing submission. In other words, a traceability table can help to identify where each method supporting each modification may be found in the marketing submission.

Table 1. Example of Description of Modifications to Modification Protocol Traceability Table

Modification Protocol Component		
Modification	Performance Evaluation Methods	Update Procedures
<i>Modification #1</i>	<i>Method A (see Section X.A)</i>	<i>Method J (see Section X.J)</i>
<i>Modification #2</i>	<i>Method A (see Section X.A)</i>	<i>Method K (see Section X.K)</i>
<i>Modification #3</i>	<i>Method B (see Section X.B)</i>	<i>Method L (see Section X.L)</i>

D. Impact Assessment

An Impact Assessment in a PCCP is the documentation of the assessment of the benefits and risks of implementing a PCCP for a device, as well as the mitigations of those risks. The manufacturer’s existing quality system should be used as the framework in which to conduct an Impact Assessment for the modifications set forth in the PCCP.

Documentation for an Impact Assessment provided to FDA in a marketing submission containing a PCCP should:

- 1) Compare the version of the device with each modification implemented individually to the version of the device without any modifications implemented;

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1082 2) Discuss the benefits and risks, including risks of harm,¹¹⁰ of each individual
1083 modification;
1084 3) Discuss how the verification and validation activities proposed within the Modification
1085 Protocol continue to reasonably ensure the safety and effectiveness of the device;
1086 4) Discuss how the implementation of one modification impacts the implementation of
1087 another; and
1088 5) Describe the cumulative impact of implementing all modifications.
1089

1090 FDA believes it is important to address these elements in an Impact Assessment in order to
1091 demonstrate that the combination of the proposed modifications is unlikely to introduce
1092 additional, unmitigated risks, and that the safety and effectiveness of the device is maintained as
1093 modifications are implemented.
1094

1095 Impact Assessment documentation for a PCCP in a marketing submission should also discuss
1096 how the individual modifications included in the PCCP impact not only the particular device
1097 function, but the overall functionality of the device, including how they impact other device
1098 software functions and/or device hardware. For combination products, such documentation
1099 should also discuss how the individual modifications included in the PCCP for the device
1100 constituent part impact the biologic and/or drug constituent part, and the combination product as
1101 a whole. Additionally, if the modifications in a PCCP are intended to affect any device functions
1102 of a multiple function device product, we recommend considering FDA’s guidance “[Multiple
1103 Function Device Products: Policy and Considerations](#).” In particular, as it pertains to the device
1104 and the PCCP, it is important to determine if any information should be included in the Impact
1105 Assessment in a PCCP in a marketing submission so that FDA may assess the impact of the
1106 “other function(s)” on the safety or effectiveness of the device as it is modified consistent with
1107 the PCCP.
1108

1109 Some information related to the Impact Assessment may be included elsewhere in your
1110 marketing submission, for example in sections for the risk assessment for the device or the
1111 Modification Protocol in your PCCP. As such, FDA recommends providing clear references in
1112 your Impact Assessment to the relevant sections in your marketing submission that support the
1113 Impact Assessment.
1114

1115 **VIII. Examples of Modifications for PCCPs**

1116 This section includes illustrative examples of modifications that generally may be or are not
1117 appropriate for inclusion in a PCCP for a specific device. Each example begins with a brief
1118 description of a device and a statement about a proposed modification for inclusion in a PCCP.
1119 Please note that the provided summaries of the devices and modifications in this section are not
1120 intended to reflect the complete content or detail expected in a Description of Modifications
1121 section in a PCCP. Rather, proposed modifications should be described in much greater detail in
1122 a PCCP, consistent with the recommendations provided throughout this guidance. Due to the

¹¹⁰ See, e.g., harm, as defined in ISO 14971 *Medical devices – Application of risk management to medical devices*, is the physical injury or damage to the health of people.

Contains Nonbinding Recommendations

Draft – Not for Implementation

1123 complexity of devices and PCCPs, it is not practical to describe all relevant considerations, or a
1124 complete PCCP, for the limited examples presented below.

1125
1126 As previously stated, recognizing there is a spectrum of risks for devices, for purposes of
1127 reviewing a PCCP, FDA intends to, among other considerations, take into account the Guiding
1128 Principles recommended in this guidance (Section IV.), the benefit-risk profile of the specific
1129 device that is the subject of the PCCP, the specific modifications being proposed, and FDA's
1130 experience applying this policy across different device types. As such, certain modifications that
1131 may be appropriate for inclusion in a PCCP for one device may not be appropriate for inclusion
1132 in a PCCP for another device. Some modifications may not be appropriate for inclusion within a
1133 PCCP for any device.

1134
1135 FDA encourages manufacturers to leverage the [Q-Submission Program](#) to obtain FDA feedback
1136 on a proposed PCCP for a device and submission type prior to submitting a marketing
1137 submission.

1138 1139 **Example 1**

1140
1141 This device is a microarray-based IVD cancer predisposition risk assessment system for over-
1142 the-counter use that is intended to detect 5 single nucleotide variants in BRCA1 and 3 single
1143 nucleotide variants in BRCA2 in saliva for the purposes of describing if a person is at increased
1144 risk of breast cancer, ovarian cancer, or prostate cancer.

1145
1146 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1147 • Addition of new single nucleotide variants
- 1148 • Addition of insertion and deletion variants up to 20 base pairs
- 1149 • Updates to the labeling that inform potential cross-reactive polymorphisms

1150
1151 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1152 • Addition of copy number variants
- 1153 • Addition of a new gene
- 1154 • Change in the collection device and sample type (e.g., saliva to buccal swab)
- 1155 • Change from manual to automated process (e.g., for sample dilution)

1156 1157 **Example 2**

1158
1159 This device is an ion selective electrode IVD intended for use on a laboratory-based chemistry
1160 analyzer to quantify the concentrations of potassium ions in serum samples for the purposes of
1161 monitoring electrolyte balance in the diagnosis and treatment of diseases and conditions
1162 characterized by low or high blood potassium levels.

1163
1164 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1165 • Addition of lithium heparin plasma as a sample type
- 1166 • Extension of sample stability claims (e.g., 2 hours at room temperature to 4 hours at room
1167 temperature)

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1168 • Addition of a new potassium ion selective electrode

1169

1170 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1171 • Addition of urine or capillary whole blood as a sample type

- 1172 • Addition of at-home sample collection

- 1173 • Addition of point of care use

1174

Example 3

1175

1176 This device is a non-absorbable polyethylene surgical suture intended for soft tissue
1177 approximation or ligation.

1178

1179 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1181 • Change to a different non-novel sterilization method (e.g., Established Category A to
1182 Established Category B sterilization method¹¹¹)

- 1183 • Extend shelf life using a different well-established method than what was provided for
1184 the initial shelf life testing

- 1185 • Addition of sutures to the product line with different dimensions that are within the range
1186 of dimensions of those currently authorized

- 1187 • Addition of dye with an appropriate FDA listed color additive per 21 CFR Part 74
1188 Subpart D¹¹²

1189

1190 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1191 • Addition of antimicrobials

- 1192 • Change in filament design to an atypical design (e.g., unique braiding patterns, anchors,
1193 or knots)

- 1194 • Addition of a stiffening agent to the ends of the suture to address a recall

1195

Example 4

1196

1197 This device is a multi-parameter physiological patient monitor with arrhythmia detection and
1198 alarms for use in a hospital environment.

1199

1200 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1201 • Hardware and software updates to introduce compatibility with a newly cleared
1202 monitoring parameter/module

- 1203 • Change to a new wireless card that has different technical specifications than those for
1204 the authorized device

1205

¹¹¹ See FDA's guidance "[Submission and Review of Sterility Information in Premarket Notification \(510\(k\)\) Submissions for Devices Labeled as Sterile.](#)"

¹¹² For additional information on color additives, see FDA's website on "[Summary of Color Additives for Use in the United States in Foods, Drugs, Cosmetics, and Medical Devices.](#)"

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1206 • Change to the software to upgrade to a new version of the currently supported operating
1207 system (OS) on the basis of a risk assessment that shows changes between major OS
1208 versions could significantly impact the software’s performance
- 1209 • Change to the visualization of patient monitoring data (e.g., heart rate) in the user
1210 interface
- 1211 • Addition of new parameter trend graphs for use in active patient monitoring
1212

1213 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1214 • Addition of a new physiological parameter to be monitored (e.g., blood oxygen level,
1215 temperature)
- 1216 • Addition of a novel physiological or predictive index or algorithm (e.g., new algorithm to
1217 predict risk of patient deterioration)
- 1218 • Significant changes to the alarm architecture
- 1219 • Significant changes to the design of the printed circuit board
1220

Example 5

1221 This device is an over-the-counter mobile medical app intended to assess risk of moderate to
1222 severe obstructive sleep apnea.
1223
1224
1225

1226 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1227 • Addition of new connected data sources to provide the same types of inputs needed by
1228 the app
- 1229 • Addition of support for mobile platforms with different operating systems, where app
1230 performance specifications have been retained (e.g., introducing an iOS version of an
1231 Android app)
- 1232 • Change in software algorithm to improve device performance by reducing false positive
1233 outputs in typical use based on additional real-world use data
1234

1235 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1236 • Addition of an alternate type of input data on which the app will make or refine its
1237 assessment of sleep apnea risk
- 1238 • Significant change to software architecture
- 1239 • Change in labeling to include claims about helping users to assess risk of mild obstructive
1240 sleep apnea
1241

Example 6

1242 This device is an IVD intended for use in determining quantitative susceptibility of Candida
1243 species to Caspofungin.
1244
1245
1246

1247 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1248 • Plans for future updates to breakpoints that are consistent with FDA’s website on [FDA-
1249 Recognized Antimicrobial Susceptibility Test Interpretive Criteria](#), consistent with the

Contains Nonbinding Recommendations

Draft – Not for Implementation

1250 recommendations in FDA’s guidance “[Antimicrobial Susceptibility Test \(AST\) System](#)
1251 [Devices – Updating Breakpoints in Device Labeling](#)”
1252

1253 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1254 • Addition of a new drug to test *Candida* species
- 1255 • Addition of claimed organism species not included in the original authorized device
- 1256 • Addition of alternative reading method (e.g., from manual (visual/turbidimetric or
1257 colorimetric) to automated (fluorescence))
- 1258 • Addition of alternative inoculation method (e.g., from manual to automated)
- 1259

Example 7

1260 This device is a human leukocyte antigen (HLA) molecular typing assay intended to aid donor
1261 and recipient matching in transplantation and transfusion. It uses polymerase chain reaction
1262 sequence-specific primer (PCR-SSP) or sequence-specific oligonucleotide probes (SSOP) to
1263 detect HLA-B alleles.
1264
1265

1266 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1267 • Addition of new primers or probes to detect new HLA-B alleles
- 1268 • Updates to the data analysis software to resolve or inform HLA typing ambiguities based
1269 on an internationally recognized HLA sequence database
- 1270

1271 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1272 • Addition of new primers or probes to detect alleles of a different HLA gene or locus
- 1273 • Change in the indications for use to include a companion diagnostic claim to identify
1274 patients who have specific HLA allele(s) and may benefit from treatment with a
1275 corresponding therapeutic product or are likely to be at increased risk for serious adverse
1276 reactions as a result of treatment with a corresponding therapeutic product
- 1277

Example 8

1278 This device is an implantable pulse generator pacemaker.
1279

1280 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1281 • Addition of an alternate component supplier (e.g., memory chip, resistor, capacitor)
1282 where the component specifications and design requirements are identical to those of the
1283 currently approved component
- 1284 • Minor software changes to improve the battery longevity estimation algorithm
- 1285 • Minor change to the wireless modem in the pacemaker to expand the range of cellular
1286 frequencies and bands
- 1287 • An update to change the Magnetic Resonance (MR) labeling for a particular device
1288 model from MR Unsafe to MR Conditional based on well-established test methods and
1289 acceptance criteria
- 1290
- 1291
- 1292
- 1293

Contains Nonbinding Recommendations

Draft – Not for Implementation

1294 Modifications that are generally not appropriate for inclusion in a PCCP:

- 1295 • A manufacturing change to an adhesive application process step made in response to
- 1296 reported device failure events of premature battery depletion due to an identified process
- 1297 variation
- 1298 • Addition of a new battery design or change to the battery chemistry
- 1299 • Addition of a new software feature to optimize therapy delivery by customizing pacing
- 1300 parameters
- 1301 • Change to device software to address reports of failure to deliver pacing therapy due to
- 1302 sensing errors
- 1303 • Change to the indications for use to include conduction system pacing
- 1304

Example 9

1305
1306
1307 This device is an immunoassay-based IVD intended to quantitatively measure prostate specific
1308 antigen in serum to aid in the detection of prostate cancer in conjunction with a digital rectal
1309 exam.

1310
1311 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1312 • Addition of lithium heparin plasma as a sample type
- 1313 • A protocol for extension of sample stability claims
- 1314 • Drug interference for the alkaline phosphatase detection mechanism that does not require
- 1315 a design change to address the interference
- 1316 • Addition of an alternate antibody supplier where the specifications for the antibody are
- 1317 identical to those of the currently approved antibody
- 1318 • In-processing manufacturing change (e.g., reagent pooling) to add a filling suite within
- 1319 the same building
- 1320

1321 Modifications that generally would not be appropriate for inclusion in a PCCP:

- 1322 • Addition of capillary whole blood as a sample type
- 1323 • Addition of point-of-care use
- 1324 • Change to remove use of the test in conjunction with a digital rectal exam
- 1325

Example 10

1326
1327
1328 This device is a coronary drug-eluting stent system, which is a device-led combination product.
1329 The stent is coated with a sirolimus/polymer blend intended to inhibit restenosis.

1330
1331 Modifications that generally may be appropriate for inclusion in a PCCP:

- 1332 • Addition of an alternate delivery system component supplier (e.g., hypotube, adhesive)
- 1333 where the component specifications and design requirements are identical to those of the
- 1334 currently approved component
- 1335 • Addition of stent systems to the product line with different dimensions, but otherwise
- 1336 identical designs, that are within the range of dimensions of those currently authorized,
- 1337 which uses identical manufacturing processes

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 1338 • Automation of a dimension measurement inspection process step in manufacturing

1339

1340 Modifications that generally would not be appropriate for inclusion in a PCCP:

- 1341 • Change in the composition of the drug coating (e.g., a change from sirolimus to a
1342 different drug, a change in the concentration or total dose of sirolimus, or a change in the
1343 ratio of polymer to sirolimus)
- 1344 • A manufacturing change that impacts the drug coating (e.g., a change in the materials or
1345 methods used to filter the coating solution)
- 1346 • Change in the labeling to increase the post-dilation expansion limit of the stent
- 1347 • Change in the labeling to include new indications, instructions, guidance, or clinical
1348 information regarding use in a new patient population (e.g., adding the results of a
1349 clinical study of the device’s use in patients with complex coronary lesions)

1350

1351 **IX. Sample of 510(k) Summary Information Regarding the** 1352 **PCCP**

1353 This section includes a sample of 510(k) summary information regarding a PCCP for a device
1354 that was authorized with a PCCP. It is based on some of the modifications included in Example 4
1355 in Section VIII. of this guidance. For complete information on the 510(k) summary, see 21 CFR
1356 807.92, and FDA’s [“The 510\(k\) Program: Evaluating Substantial Equivalence in Premarket](#)
1357 [Notifications \[510\(k\)\]”](#) guidance, including Appendices B. and C.

1358

1359 **Predetermined change control plan**

1360

1361 The predetermined change control plan (PCCP) for the device specifies anticipated modifications
1362 to the device software to add a new wireless card and to upgrade the device’s operating system.
1363 The PCCP also specifies the methods to implement those modifications so that the device
1364 remains as safe and as effective as the predicate device. The detailed description of the
1365 modifications, testing methods, validation activities, performance requirements, and
1366 communication to users are summarized in the table below.

1367

Contains Nonbinding Recommendations

Draft – Not for Implementation

Planned Modifications	Test Methods and Validation Activities	Communication to users, as needed
<p>Addition of a new wireless card</p>	<p>Testing for basic safety and essential performance will be repeated according to FDA-recognized editions of the following voluntary consensus standards:</p> <ul style="list-style-type: none"> • ANSI/AAMI ES60601-1 • IEC 60601-1-2 • IEC 60601-1-8 • IEC 60601-4-2 • AIM 7351731 • IEEE/ANSI C63.27 • AAMI TIR69 <p>Test methods will follow the recommendations in the most recent version of the following FDA guidances:</p> <ul style="list-style-type: none"> • “Electromagnetic Compatibility (EMC) of Medical Devices” • “Radio Frequency Wireless Technology in Medical Devices” 	<p>Labeling will be updated in accordance with the authorized PCCP to provide users with current information regarding the device’s wireless capabilities</p>
<p>Upgrade to device operating system</p>	<p>Verified and validated by the requirements of the Modification Protocol, including necessary regression testing</p>	<p>Users will be notified of software updates (including any user instructions for the update process) in accordance with the authorized PCCP</p> <p>Labeling will be updated in accordance with the authorized PCCP to provide users with current information regarding the device’s upgraded operating system</p>